



INTRODUCTION

Trulife's Aluminum Adapters are designed for use in endoskeletal construction of Trulife limb systems. It provides $\pm 8^\circ$ angular adjustment. For maximum allowable patient weights, please refer to the chart below.

Product Code	Description	Weight Limit
AAA208	Double Offset Four Hole Pyramid Adapter AL	160 kg / (353 lb) (Med. Activity)
AAA209	Single Offset Four Hole Pyramid Adapter AL	160 kg / (353 lb) (Med. Activity)
AAA210	Four Hole Pyramid Adapter AL	160 kg / (353 lb) (Med. Activity)
AAA211	Four Hole Female Adapter AL	160 kg / (353 lb) (Med. Activity)

INSTALLATION AND USE

Recommended installation and use procedures must be followed for maximum safety and service life.

Warning: Never modify any Trulife Aluminum Adapter unless directed to in this installation guide. Any modifications to an adapter will void the warranty and can lead to failure.

Angular Alignment Instructions for all Trulife's Aluminum Adapters

1. Place the mating male pyramid into the female receiver end and tighten the set screws while maintaining appropriate alignment.
2. Alternately loosen and tighten opposing set screws until the desired alignment is achieved. Make sure all set screws are tight before allowing the patient to bear weight on the limb.

To maintain integrity of the assembled prosthesis, perform these final assembly steps:

3. Apply a generous amount of [BLUE] Loctite 242 (or equivalent) removable thread retaining compound to the threads of all fasteners.
4. Refasten the set screws and tighten to 15 Nm (11 ft-lbs/132 in-lbs) using a 4 mm hex driver and torque wrench.
5. Allow patient to load and/or walk the prosthesis.
6. Re-apply 15 Nm torque to all set screws.

Important: Follow steps 5 and 6 above to account for any settling or compression set of the assembled components.

- Final application of fastener tightening torque must be done within 2 hours of initial application of thread retaining compound.
- DO NOT use high-strength forms of thread retaining compound (example: Red or Green Loctite formulations).
- Insufficient fastener torque will lead to fatigue failure of any modular adapter.
- Excessive fastener torque can cause component failure; stripped threads and/ or stress fractures.

**Four-Hole Adapters (AAA208, AAA209, AAA210, AAA211)**

The 4-Hole Adapters will connect to components that have a corresponding 4-hole, 2 inch diameter bolt circle configuration, with holes tapped to thread size M6x1.0.

AAA208 is a double offset male pyramid, ↑5mm, →5mm from center of 4-hole pattern.

AAA209 is a single offset male pyramid, ↑7mm from center of 4-hole pattern.

AAA210 is a male pyramid adapter with the pyramid located at the center of the 4-hole pattern.

AAA211 is a female adapter centered on the 4-hole pattern.

1. Apply a generous amount of [BLUE] Loctite 242 (or equivalent) removable thread retaining compound to the four mounting screws provided (size M6 x 1.0 x 30mm long).
2. Assemble the adapter to the mating component and evenly tighten the screws using a 4mm hex driver.
3. In a cross-pattern tighten the screws to 9.5 Nm (7 ft-lbs/ 84 in-lbs) using a torque wrench.
4. Allow patient to load and/or walk the prosthesis.
5. Re-apply 9.5 Nm torque to all screws

For all Trulife's Aluminum Adapters

1. Apply a generous amount of [BLUE] Loctite 242 (or equivalent) removable thread retaining compound to the threads of all fasteners.
 2. Apply the specified torque to the fastener while the thread retaining compound is still wet.
 3. Allow patient to load and/or walk the prosthesis.
 4. Re-apply torque to all fasteners within 2 hrs of initial application of thread retaining compound.
- Use only bolts, set screws and other fasteners supplied or recommended by Trulife. Use of unapproved fasteners will void the warranty and may lead to component failure.
 - When adjusting alignment or re-assembling components that had been previously assembled with thread retaining compounds, the threads of the fasteners and threaded holes should be cleaned and free of any retaining compound residue to ensure a free-running thread fit. Use a mild solvent such as alcohol.

Warning: Failure to follow the installation and use procedures set forth above may lead to structural failure of the components subjecting the user to a risk of serious personal injury.

Warning: Torque settings should be checked periodically. A loose fastener may lead to component failure.

Warning: Stop use immediately if a fastener needs to be re-tightened more than twice. Repeated re-tightening may lead to component failure.

Warning: Never reuse bolts. If you need to reinstall Trulife components for any reason, please call Trulife for new fasteners.

Warning: Do not contaminate fasteners with any type of paint, glue, or cement, except for the recommended



thread retaining compounds. For all fasteners, ensure a free-running thread fit prior to applying final tightening torque.

QUESTIONS

Contact Customer Service in the U.S. at 800-492-1088, or fax 800-245-3765

If calling from outside the U.S., contact Customer Service at 517-787-1600, or fax 360-697-6843

Visit Trulife online at www.trulife.com

LIMITED WARRANTY

Trulife warrants that the Aluminum Adapters will be free from defects in material and workmanship for two (2) years from the date of installation.

This warranty will not apply if the product has been damaged by misuse, abuse, neglect, improper care, failure to follow instructions, abnormal wear and tear, or in the event that the Aluminum Adapters have been modified/repared by persons unauthorized by Trulife.

If a defect in material or workmanship is found during the warranty period, Trulife will, at Trulife's option, either repair or replace the product. If it is not possible to repair or replace the product, Trulife will be limited to refunding the purchase price.

Trulife will not be liable under any legal theory for any direct, indirect, special, incidental or consequential damages arising from the use of or inability to use this product.

The application guidelines for this Trulife product are for the use of and by certified, qualified practitioner only. Patients are not to attempt to apply or adjust the item unless expressly instructed to do so by the practitioner responsible for the prescription and/or initial fitting of the device. All patient questions should be referred to the practitioner and not to the manufacturer. The manufacturer warrants only that the enclosed product has been inspected for quality and can be effective for certain indications, but final decisions and ongoing monitoring must be made by the prosthetic professional(s) prescribing and/or fitting the device to determine its effectiveness for an individual patient. Patient compliance is an integral part of the entire protocol and must be adhered to in order to avoid potential problems and to maximize the effectiveness of the prescribed product.

As a condition of the sale of any Trulife product, this prosthesis is restricted to a "Single Patient Use Only" by the originally fitted patient in order to protect the care provider and the patient against potentially adverse consequences of infectious disease transmission, material instability in adapting to the configuration of the original user and/or decrease in effectivity. Any express or implied warranties are voided if the product is reused or fitted to another patient. Additionally, a license of right to use under any relevant patents pertaining to the product is terminated with the cessation of use by the original patient. As with all Trulife products, this product must be prescribed and applied by a qualified practitioner to determine it meets the needs of the particular patient and accomplishes the desired results.